



Please note that before introducing the tube into the trocar cannula it is necessary that you make sure to stretch the bag about 5mm thus that it loosens from the white tube

STERILE EO

NF EN ISO 13485
ISO 9001

LiNA Medical ApS
 Formervangen 5
 DK-2600 Glostrup
 DENMARK
 Tel: +45 43 29 66 66 – Fax: +45 43 29 66 99
 Email: linamed@lina-medical.com
 www.lina-medical.com